


TEST KIT CONTENTS



Sterile swab (1x)




Tube with diluent (1x)



Dropper with protective cap (1x)



Plastic Biohazard Disposal Bag



Instructions for Use (1x)



Test cassette (1x)


HOW TO USE

1 – Anterior nasal sample collection.

- Remove the sterile swab from the peel-pack being careful to avoid touching the sponge tip with your hands or contacting it with any foreign surface.
- Insert the swab sponge tip 1.5cm (3/4 inch) into your nostril.
- Slowly rotate the swab in a circular path against the inside of your nostril at least 4 times for a total of 15 seconds.
- Using the same swab, repeat steps (B) and (C) in the other nostril.



Left nostril - 15 seconds



Right nostril - 15 seconds

2 – Preparing the sample specimen.

- Fully immerse the swab sponge tip into the diluent. To ensure all the sample from the swab is mixed into the inserting diluent, rotate the swab stick in a twisting motion in the tube diluent for 30 seconds while squeezing tube against the swab sponge tip.



- Continue to squeeze the sides of the tube while removing the swab so as to extract all the liquid from the swab sponge tip. Then safely dispose of the swab.



- Hold the dropper by the protective cap while long-pointed end into the tube until it is securely fitted.




3 – Conducting the test.

- Gently mix the sample again by 15-30 rotating the tube for 5 seconds stirring motion. Then remove the protective cap from the dropper.
- Remove the test cassette from its foil pack (if you haven't previously done so).




- Invert the tube and squeeze 3 drops into the smaller rectangular sample well at the bottom of the test cassette (not the larger rectangular C/T results window at the top of the cassette). **Note: you may continue the test even if you inadvertently applied 4 drops.**




- Start the stopwatch and read the test results after 15 minutes.

Note: after 30 minutes, test results no longer considered valid.



4 – Interpreting test result.


Negative test result



If the "T" line is blank and the "C" line displays a visible colour band (even if faint) then the result is negative and you are unlikely to have COVID-19.

Note: Negative results are presumptive should not be used as the sole basis for infection control decisions, particularly in the presence of clinical signs and symptoms consistent with COVID-19, or in those who have been in close contact with the virus. In these cases, it is recommended that the test is repeated after 1-2 days. In all cases, a definitive diagnosis must be made by a doctor and local health and safety control measures should always be adhered to.


Positive test result



If both the "T" line and "C" line display visible colour bands (even if faint), then the test result is positive and you most likely are infected with COVID-19.

Note: Positive results require strict adherence to your local public health authority's guidelines for suspected infections - including confirmatory PCR screening at your local hospital or testing station, self-isolation, and reporting to the applicable authority always be adhered to.

Invalid test result



If the "C" line is blank, then the test result is considered invalid.

Note: Invalid results are represented by a complete absence of a visible "C" line. In this case, carefully read the "Instructions for Use" and repeat the test. If a repeated test result is invalid, then contact your local product supplier.